APPENDIX N BULK MILK TANKER SCREENING TEST FORM

CHARM® SL (Raw Commingled Cow, Sheep, Water Buffalo and Goat Milk), IMS #9-C13 AND

Charm 3 SL3 (Raw Commingled Cow Milk), IMS #9-C15 AND

Charm BL30SEC (raw commingled Cow milk), IMS #9-C20

BETA-LACTAM TESTS

FDA/NCIMS Revision 10/19

[Unless otherwise stated all tolerances are ±5%]

GENERAL REQUIREMENTS

1.	See Appendix N General Requirements (App. N GR) items 1-8 & 15						
				SAMPLES			
2.	See	Арр	. N G	R item 9			
				APPARATUS & REAGENTS			
3.	. Equipment						
	a.	Charm Sciences Strip Incubator 56±1°C 8 min timer – SL beta-lactam test; 56±1°C 3 min with internal timer – SL3 beta-lactam test; 56±1°C Charm EZ display when message "Add milk to strip and close door" Required for BL30SEC optional for SL and SL3					
		1.	Clea	an, properly maintained and located on a level surface			
		2.	Che	eck temperature daily (day of use); maintain records			
			a.	Charm EZ printout acceptable for daily temperature check (annual accuracy check required); maintain records			
		3.	Tem	nperature measuring device for each incubator (App N. GR item 3)			
		4.	Lid	closed (slightly sprung so that timer not active) when not running tests			
		5.	Incu	ubator Temperature:			
		6.		er if not included in incubator ubation Time of internal timer:			

b.	Cha	SA® Reader, ROSA Pearl Reader (with or without ROSA Barcode option), arm EZ Reader or Charm Sciences equivalent with print out or download of a; manual available							
		Serial Number:							
	1.	ROSA Reader V1.03 or higher							
		a. Calibrators							
		Range(s)	Result						
		Low:							
		High:							
		b. Maintain records	_						
	2.	ROSA Pearl Reader V3.00 or higher	_						
		a. Calibrators - Low and High for use in all assay channels							
		Range(s)	Result						
		Low:							
		Low:(darker magenta)							
		High: (lighter pink)							
		b. Maintain records							
	3.	Charm EZ reader							
		a. Calibrators - Low and High for use in all assay channels							
		Range(s)	Result						
		,							
		Low: (darker magenta)							
		High:(lighter pink)							
		b. Maintain records							
	4.	Calibrator serial numbers match reader SN							
	5.	Do not proceed if out of range. Manufactur corrective actions	er should be contacted for						
	6.	Printer or computer link for hardcopy downloa	nd.						

	C.	Pipettor - 300 μL and disposable tips (see App. N GR item 7)					
	d.	Or single use 300 µL ROSA-pipet with overflow bulb to accurately measure amount of sample, supplied by manufacturer (screening only)					
	e.	Optional Centrifuge (Not applicable to SL3 or BL30SEC beta-lactam test) – mini or equivalent (1200-2000 x g) for frozen controls					
4.	Rea	agents					
	a.	Test Strips (EZ Compatible for Charm EZ)					
		Lot # Exp. Date:					
		QC Date: By:					
	b.	Positive Control					
		Lyophilized or tablet 5 ppb Penicillin G beta-lactam tests					
		Lot #: Exp. Date:					
	c.	Negative Control					
		Previously negative tested raw milk (item 5.d)					
5.	Reagent Stability						
	a.	SL3 and BL30SEC reagents must be received within 72 hours if shipped non-refrigerated; over 72 hours must be refrigerated. (Not applicable to the SL reagents)					
	b.	Store reagents at 0.0-4.5°C, desiccant blue, maintain no longer than manufacturer's expiration date					
		Do not use if desiccant indicator is white or pink					
	C.	Positive Control - Manufacturer supplied, maintain no longer than manufacturer's expiration date					
		Reconstitute with Negative Control (raw milk), tested +400 or more positive, used within 48 hours when maintained at 0.0-4.5°C					
		Lab Prep. Date: Lab Exp. Date:					
		 Or, aliquot within 24 hours and freeze at -15°C or colder in a non-frost-free freezer or in an insulated foam container in a frost-free freezer; use within 2 months 					
		Lab Prep. Date: Lab Exp. Date:					

		а	Mix well until sample is homogeneous				
			1. Do not use if there is visible protein precipitation				
		b.	Store at 0.0-4.5°C and use within 24 hours; do not refreeze				
			For SL ONLY , centrifuge 3 min and cool				
			Test portion below fat layer without mixing				
	3.	Day	of use, must produce +400 or greater reading; maintain records				
		Tes	t Value:				
		Do	not proceed if out of range				
d.	_		Control - raw milk tested -600 or more negative (SL Test Negative can be any of the approved species milk)				
	Sample ID: Test Value:						
	Date tested:						
	1.	I. Use within 72 hours when maintained at 0.0-4.5°C					
	2.	Or, aliquot within 24 hours and freeze at -15°C or colder in a non-frost-free freezer or in an insulated foam container in a frost-free freezer; use within 2 months					
		Lab	Prep. Date: Lab Exp. Date:				
		а	Thaw slowly overnight in refrigerator or more rapidly in cold water. Mix well until sample is homogeneous				
			Do not use if there is visible protein precipitation				
		b.	Store at 0.0-4.5°C and use within 24 hours; do not refreeze				
		C.	For SL ONLY , centrifuge 3 min and cool				
			Test portion below fat layer without mixing				
	3.	Day	of use must produce -600 or more negative; maintain records				
		Do	not proceed if out of range				

TECHNIQUE

ô.	Daily Performance and Operation Check						
	a.	See App. N GR items 10.b-d					
	b. c.	If using ROSA reader Versions 1.05 and higher, or ROSA Pearl, use ESC 5 reader function to enter performance monitoring mode of reader; if using Charm EZ, use Menu to enter Performance Monitoring mode and "Perf Mon" to enter daily performance check; refer to manual for directions Check Calibrators; items 3.b.1.a, 3.b.2.a, or 3.b.3.a					
	d.	Positive and negative controls must give appropriate readings prior to any sample analysis (see App. N GR item 10.b)					
	e.	Controls valid when in performance monitoring mode, ROSA reader version 1.05 and higher, ROSA Pearl or Charm EZ					
	f.	Do not proceed if c, d and e are not met					
7.	Tes	t Procedure					
	a.	Set out required number of test strips and place them in a dry labeled container at room temperature, or take out strips as needed					
		Discard unused test strips at the end of the day					
	b.	Label test strips, one for each test sample and each control. Avoid crushing sample compartment(s)					
	C.	Mix milk sample(s)/control(s) 25 times in 7 sec with a 1 ft movement or vortex for 10 sec at maximum setting; use within 3 min (samples/controls must be in appropriate containers to allow the use of vortexing)					
		Centrifuge sheep milk sample(s)/controls that have been previously frozen; refer to 5.c.2.a-c and 5.d.2.a-c					
	d.	Place strip into appropriate incubator					
		EZ reader in incubate and read mode displays appropriate test name on strip insertion and displays "add milk to strip and close door" when in correct temperature range, 55-57C, Follow item 7.i					
	e.	While holding strip flat, peel back plastic (to 'peel to here' line) to expose sample pad compartment. Avoid lifting the wick and sponge under tape					
		For multiple samples, complete steps 7.d-g for each sample/control, before starting test of next sample					
		Complete all samples within 2 min (1 min 15 sec for SL3 test) of placing first strip in incubator					

f.	Add 300 µL of mixed sample/control to corresponding strip							
	1.	Using pipettor (item 3.c) with new tip for each sample/control, draw up 300 µL avoiding foam and bubbles						
		a.	Rer	move tip from liquid				
		b		ile holding the pipettor vertically, expel test portion slowly into er side well of appropriate strip				
	2.	Using new manufacturer-provided ROSA-pipet (item 3.d) for each sample/control [Screening only]						
		a.	a. Squeeze top bulb while holding vertically with bulb and overflow reservoir side pointing down, draw up test portion avoiding foam and bubbles. Sample should completely fill pipet shaft and overflow into the bottom half of the overflow reservoir					
		b.	Rer	move tip from liquid				
		C.	eith	ile holding the ROSA-pipet vertically, expel test portion slowly into er side well of appropriate strip. Excess portion should remain in ervoir				
g.	Re-	seal p	al plastic firmly around sample pad compartment					
h.	ROSA Reader and Charm EZ (read only mode)							
	1.	Close lid and latch ROSA incubator to start automatic timer in the incubator. If no automatic timer in incubator, set external timer for 8 min for SL. For SL test, incubate 8 min not to exceed 9 min. For SL3 test, incubate 3 min not to exceed 3 min and 30 sec						
	2.	. At end of incubation visually inspect C (Control) line. An absent C line, a partial C line or an indistinct C line indicates an invalid test; and the sample/control must be re-tested						
	3.	Inse	ert on	ly valid test(s) in reader				
		a.	RO	SA reader set to appropriate channel				
			1.	SLBL slow blink for SL beta-lactam test				
			2.	SLBL solid (no blink) for SL3 beta-lactam test				
			3.	Press ENTER, reading and interpretation appear in 5 sec, read strips within 5 min (3 min with SL3) of completion of incubation. Strips may be held vertically, sample compartment down while waiting to be read				

			b.	Cha	rm EZ automatically sets channel when color coded strip inserted				
				1.	Close door; reading and interpretation appear in 5 sec; read strips within 5 min (3 min with SL3) of completion of incubation. Strips may be held vertically, sample compartment down while waiting to read				
	i.	Cha	rm E	Z (ind	cubate and read mode)				
		1.	Charm EZ automatically sets channel and incubator temperature when color coded strip inserted. Optionally enter sample ID						
		2.	Pee	l strip	(7.e), add milk (7.f), and re-seal strip (7.g)				
		3.	Clos	se do	or to begin				
		4.	Cha	rm E	Z automatically prompts for further testing when positive				
8.	Interpretation with Reader								
	a. If there is a negative or zero reading on the reader, sample is a Negative (NF)								
	b. If there is a positive reading on the reader, sample is an Initial Positive								
9.	Verification of Initial Positive Tanker Samples (see App. N GR item 11); Confirmation of Presumptive Positive Tanker Samples (see App. N GR item 12); and Traceback of Producer(s) on a Confirmed Positive Tanker (see App. N GR item 13)								
10.	Reporting (see App. N GR item 14)								