## **ELECTRONIC SOMATIC CELL COUNT**

## Fossomatic™ Minor (Raw Commingled Cow, Sheep, Goat, Water Buffalo and Camel Milk) IMS #16f

(Unless otherwise stated all tolerances ±5%)

1.	Laboratory Requirements (see Cultural Procedures (CP) items 33 & 34)				
	a.	Un-preserved samples may be tested up to 72 hours after initial collection			
	b.	Samples may be tested up to 7 days after initial collection if preserved with 0.02% 2-bromo-2-nitropropane-1,3-diol (Bronopol <sup>TM</sup> ) or 0.05% potassium dichromate ( $K_2Cr_2O_7$ )			
2.	Comparative Test with DMSCC [NOT required as a co-requisite for certification of analysts in laboratories purchasing standards from a CERTIFIED provider (item 9.b)]				
	a.	. Analyst(s) certified for DMSCC			
	<ul> <li>Each analyst seeking certification for the ESCC test shall perform the comparative test</li> </ul>				
		Test 4 samples (100K-200K, 300K-500K, 600K-800K and 900K-1.2M) in triplicate for both DMSCC (three separate smears each) and ESCC			
		Results must be evaluated by State/Federal LEO and shown to be acceptable prior to official use of test in laboratory			
		Copy of comparison and results in QC record (or easily accessible on file in the laboratory); kept for as long as analyst is certified			
	C.	Required for laboratories preparing in house standards or using commercially prepared standards (items 9.a and c) and for those testing goat or camel milk			
		APPARATUS			
3.	See	CP items 1-4			
4.	Wat	er Bath			
	a.	Circulating and thermostatically controlled to 37-42°C			

## **REAGENTS**

<b>5</b> .	Rea	agents			<del></del>		
	a.	Dye Solution	Lot #:	Exp. Date:			
	b.	Clean 1	Lot #:	_ Exp. Date:			
	C.	Clean 2	Lot #:	_ Exp. Date:			
6.	Pre	paration					
	<ul> <li>Ready to Use Dye Solution: Pour into a clean glass container designated for the Dye Solution (item 5.a). Use within 4 weeks of dispensing into container.</li> </ul>						
		Date Disp	ensed:	Exp. Date:			
<ul> <li>Clean 1 Solution: In a clean glass container, mix one unit (20 mL) Clean 1 (item 5.b) with deionized (DI) or Microbiologically Suitable (MS) water to make 1 L, store and use within 4 weeks; when stored at 2-8°C, use within 8 weeks</li> </ul>							
		Date Prep	D.:	Exp. Date:			
	c. Ready to Use Clean 2 Solution: Pour into a clean glass container designated for the Clean 2 Solution (item 5.c). Use within 4 weeks of dispensing into container.						
		Date Disp	ensed:	Exp. Date:			
7.	All	solutions labele	ed with date prep	pared and expiration date			
				START UP			
8.	Cel	l Counter					
	a.			lean 1 and Clean 2 solutions in the supply mber of samples to be tested			
	b.	Solutions not u	sed beyond expira	ation date(s)			
	C.		tart Up" Job seque e-check the zero	ence: If the Zero Count is > 6, repeat "Clean			
	d. IF ANY ABOVE PARAMETERS ARE OUT OF VARIANCE, CORRECT BEFORE PROCEEDING						
	e.	Maintain record	ds on all paramete	ers each time instrument is used			

). N	lilk Sta	andards	
а	. Co	ommercially prepared:	
		Lot#: Date Rcd:	
	1.	Four standards in ranges 100K-200K, 300K-500K, 600K-800K and 900K-1.2M	
	2.	Perform DMSCC in triplicate on each standard in set and average counts; maintain records	
	3.	Perform DMSCC check in rotation by all certified analysts	
	4.	Use standards within one week	
		Lab Exp. Date:	
b	. Ce	ertified provider:	
		Lot#: Exp. Date: Date Rcd:	
	1.	Four standards in ranges 100K-200K, 300K-500K, 600K-800K and 900K-1.2M	
	2.	Maintain copies of all provided DMSCC values	
	3.	Measure and maintain records of temperature (0.0-7.5°C) of standards as received	
	4.	Maintain copies of all correspondence regarding problems	
	5.	Standards used by manufacturer's expiration date	
	6.	Failed standards shall be verified with DMSCC	
		If no analysts certified for DMSCC then a new set of standards is required	
		b. Do not continue with official testing until the new standard(s) test(s) in range	
C	. Lal	boratory prepared (weekly)	
	1.	Prepare from raw milk > 18 hours old preserved with 0.05% potassium dichromate (K <sub>2</sub> Cr <sub>2</sub> O <sub>7</sub> )	
	2.	Or, preserved with 0.02% 2-bromo-2-nitropropane-1,3-diol (Bronopol™)	
	3.	Standards <b>cannot</b> be preserved with formalin	

		4.	Prepare 4 standards in ranges 100K-200K, 300K-500K, 600K-800K and 900K-1.2M; use within one week			
			Lab Prep. Date: Lab Exp. Date:			
		5.	Perform DMSCC in triplicate on each standard and average counts; maintain records			
		6.	Perform DMSCC check in rotation by all certified analysts			
	d.	Hourly Control Sample (instrument drift check)				
		1.	Use one of the standards (items 9.a, b or c) in the 600-800K range, test in triplicate and determine average			
		2.	Optionally, prepare sufficient control/sample 600-800K range, test in triplicate and determine average			
			PROCEDURE			
10.	Tes	esting Standards (each time instrument used)				
	a.	Heat standards to 37-42°C (using a temperature control) and test within 30 min of reaching temperature, use once and then discard, i.e. do not re-use				
	b.	Mix by inverting at least 10x, test standards within 3 min				
	C.		et the standards in triplicate and average the counts for each level; intain records			
	d.		ch standard's average must be within 10% of the DMSCC (item 9) for that el, except within 15% for 100K-200K standard; maintain records			
	e.	of v	peatability - a standard in the 300K to 800K range must have a coefficient variation (CV) of 5% or less on 10 replicates (Refer to Operating Manual); intain records			
	f.	f. Alternatively, set and run standard check as a "Custom Job", enter DMSC values (item 9) into Excel™ macro, starting the job will enable 10.c throug 10.e to be run and calculated automatically				
	g.	THE	ESE PARAMETERS MUST BE ACHIEVED BEFORE PROCEEDING			
11.	Tes	ting	Samples			
	a.		at samples to 37-42°C (using a temperature control) and test within 30 min eaching temperature			
	b.	et samples within 10 min after removal from water bath				
	C	Mix	by inverting at least 10x, test samples within 3 min			

	d.	Record number of cells counted for each sample			
12.	With	h Continuous Operation:			
	a.	Perform a zero check (item 8.d) hourly			
	b.	Test a standard or optionally a control/sample (item 9) in the 600K to 800K range hourly in triplicate and determine the average, must be within 5% of the original established instrument average value (optionally, within 10% of original DMSCC average)			
	c.	Maintain records			
13.	Rou	utine maintenance			
	a.	Maintain records			
		REPORTING			
14.	Con	omputing and Reporting Counts			
	a.	Count obtained x 1000 is the cell count/mL milk			
	b.	In reporting electronic somatic cell counts (ESCC/mL); record only first two left hand digits, raising second digit to next higher number when third digit is 6 or more			
	C.	Report the two left hand digits (rounded)			
		If the third digit is 5 the second digit is rounded by the following rule			
		a. When the second digit is odd round up, raise the second digit by 1 (odd up, 235 to 240)			
		b. When the second digit is even round down, delete the 5 and report the second digit as is (even down, 225 to 220)			
	d.	If count on instrument is < 100 report as < 100,000 ESCC/mL			
	e.	If goat or camel milk is over the regulatory limit, follow confirmation procedure in PMO			