

APPENDIX N BULK MILK TANKER SCREENING TEST FORM

**CHARM® SL (Raw Commingled Cow, Sheep, Water Buffalo and Goat Milk), IMS #9-C13
AND
CHARM 3 SL3 (Raw Commingled Cow Milk), IMS #9-C15
AND
CHARM BL30SEC (Raw Commingled Cow Milk), IMS #9-C20**

BETA-LACTAM TESTS

[Unless otherwise stated all tolerances are ±5%]

GENERAL REQUIREMENTS

- 1. See Appendix N General Requirements (App. N GR) items 1-8 & 15** _____

SAMPLES

- 2. See App. N GR item 9** _____

APPARATUS & REAGENTS

- 3. Equipment** _____

- a. Charm Sciences Strip Incubator:
56±1°C 8 min timer – SL beta-lactam test;
56±1°C 3 min with internal timer – SL3 beta-lactam test;
56±1°C Charm EZ or EZ Protect+ (required for BL30SEC optional for SL and SL3) _____

1. Clean, properly maintained and located on a level surface _____

2. Check temperature daily (day of use); maintain records _____

- a. Charm EZ and EZ Protect+ printout acceptable for daily temperature check (annual accuracy check required); maintain records _____

3. Temperature measuring device for each incubator (App N. GR item 3) _____

4. Lid when not running tests _____

- a. 4 strip incubator lid closed (slightly sprung so that timer not active) _____

b. 2 strip incubator lid may be open _____

c. Charm EZ- reader lid open _____

d. Charm EZ-Protect+ not applicable _____

5. Incubator Temperature: _____

6. Timer if not included in incubator
Incubation Time of internal timer: _____

b. ROSA ® Reader, ROSA Pearl Reader (with or without ROSA Barcode option), Charm EZ Reader, EZ-Protect+, or Charm Sciences equivalent with print out or download of data; manual available _____

Serial Number: _____

1. ROSA Reader V1.03 or higher _____

a. Calibrators _____

Range(s)	Result
Low: _____	_____
High: _____	_____

b. Maintain records _____

2. ROSA Pearl Reader V3.00 or higher _____

a. Calibrators - Low and High for use in all assay channels _____

Range(s)	Result
Low: _____ (darker magenta)	_____
High: _____ (lighter pink)	_____

b. Maintain records _____

3. Charm EZ reader _____

a. Calibrators - Low and High for use in all assay channels _____

Range(s)	Result
Low: _____ (darker magenta)	_____
High: _____ (lighter pink)	_____

b. Maintain records _____

4. Charm EZ-Protect+ each optics module in base unit _____

a. Calibrators – Low and High for use in all assay channels. Shall be tested in all optics modules used each day attached to EZ Protect+ base with matching serial number _____

b. Low and High Calibrators results – IN RANGE=1, OUT OF RANGE=0
Result(s) _____

Low: _____
(darker magenta)

High: _____
(lighter pink)

c. Maintain records _____

5. Calibrator serial numbers match reader SN or base of EZ Protect+ _____

6. **Do not proceed if out of range.** Manufacturer should be contacted for corrective actions _____

7. Printer or computer link for hardcopy download _____

c. Pipettor - 300 µL and disposable tips (see App. N GR item 7) _____

d. Or single use 300 µL ROSA-pipet with overflow bulb to accurately measure amount of sample, supplied by manufacturer (**screening only**) _____

e. Optional Centrifuge (Not applicable to SL3 or BL30SEC beta-lactam test) – mini or equivalent (1200-2000 x g) for frozen controls _____

4. Reagents _____

a. Test Strips _____

Lot #: _____ Exp. Date: _____

QC Date: _____ By: _____

b. Positive Control _____

1. Lyophilized or tablet 5 ppb Penicillin G beta-lactam tests _____

Lot #: _____ Exp. Date: _____

c. Negative Control _____

1. Previously negative tested raw milk (item 5.e) _____

d. Negative Control Qualifier (If needed) _____

1. Freeze dried Negative Control Qualifier _____

Lot #: _____

Exp. Date: _____

5. Reagent Stability

- a. SL3 reagents must be received within 7 days (168 hours) if shipped non-refrigerated; over 7 days must be refrigerated. (Not applicable to the SL reagents)
- b. BL30SEC reagents must be received within 72 hours if shipped non-refrigerated; over 72 hours must be refrigerated. (Not applicable to the SL reagents)
- c. Store reagents at 0.0-4.5°C, desiccant blue, maintain no longer than manufacturer's expiration date

1. Do not use if desiccant indicator is white, pink or purple

- d. Positive Control - Manufacturer supplied, maintain no longer than manufacturer's expiration date
 - 1. Reconstitute with Negative Control (qualified raw milk item 5.e), tested +400 or more positive, used within 48 hours when maintained at 0.0-4.5°C

Lab Prep. Date: _____ Lab Exp. Date: _____

- 2. Or, aliquot within 24 hours and freeze at -15°C or colder in a non frost-free freezer or in an insulated foam container in a frost-free freezer; use within 2 months

Lab Prep. Date: _____ Lab Exp. Date: _____

- a. Thaw slowly overnight in refrigerator or more rapidly in cold water. Mix well until sample is homogeneous

1. Do not use if there is visible protein precipitation

- b. Store at 0.0-4.5°C and use within 24 hours; do not refreeze

- c. For **SL ONLY**, centrifuge 3 min and cool

- 1. Test portion below fat layer without mixing

- 3. Day of use, must produce +400 or greater reading; maintain records

Test Value: _____

Do not proceed if invalid

- e. Negative Control - raw milk tested -600 or more negative; (SL Test Negative Control can be any of the approved species milk)

Sample ID: _____ Test Value: _____

Date tested: _____

1. Use within 72 hours when maintained at 0.0-4.5°C
2. Or, aliquot within 24 hours and freeze at -15°C or colder in a non-frost-free freezer or in an insulated foam container in a frost-free freezer; use within 2 months

Lab Prep. Date: _____ Lab Exp. Date: _____

- a Thaw slowly overnight in refrigerator or more rapidly in cold water. Mix well until sample is homogeneous

1. **Do not use if there is visible protein precipitation**

- b. Store at 0.0-4.5°C and use within 24 hours; do not refreeze

- c. For **SL ONLY**, centrifuge 3 min and cool

1. Test portion below fat layer without mixing

3. Day of use must produce -600 or more negative; maintain records

Do not proceed if invalid

- f. Optional Use of Negative Control Qualifier (NCQ) - Negative Control Qualifier is used if no previously tested Negative Control (qualified raw milk, item 5.e) is available

1. Rehydrate NCQ

- a. Allow bottle to warm at room temperature for 10 minutes
- b. Reconstitute with 5.0 mL of 45°C potable water. Shake well.
- c. Allow to stand refrigerated or on ice for 15 minutes.
- d. Shake before use
- e. Use within 48 hours when maintained at 0.0-4.5°C
- f. DO NOT use NCQ to rehydrate Positive Control

2. Qualification of raw milk as Negative Control

- a. Test prepared NCQ twice
 1. Both NCQ readings must be -600 or more negative
- b. Test raw commingled milk twice

1. Both raw commingled milk readings must be -600 or more negative _____
- c. Average the two NCQ readings and the two raw commingled milk readings _____
 1. The average of the two commingled milk readings must be within ± 1300 of the average of the two NCQ readings _____
 2. If these conditions are met, the raw commingled milk is qualified as a Negative Control, item 5.e _____
- d. Proceed to item 6 to run Performance Monitoring _____

TECHNIQUE

6. Daily Performance and Operation Check _____

- a. See App. N GR items 10.b-d. If a reader/optics module is used to test multiple tests, one of those tests shall be performance tested daily _____
- b. If using ROSA reader Versions 1.05 and higher, or ROSA Pearl, use ESC 5 reader function to enter performance monitoring mode of reader; if using Charm EZ, use Menu to enter Performance Monitoring mode and "Perf Mon" to enter daily performance check; if using Charm EZ Protect+, use "PM" button and select "Performance Monitoring" to enter daily performance check. Each EZ Protect+ optics module in use must complete Daily Performance and Operation Check, item 6; refer to manual for directions _____
- c. Check Calibrators; items 3.b.1.a, 3.b.2.a, 3.b.3.a, or 3.b.4.a _____
- d. Positive and Negative Controls must give appropriate readings prior to any sample analysis (see App. N GR item 10.b) _____
- e. Controls valid when in performance monitoring mode, ROSA reader version 1.05 and higher, ROSA Pearl or Charm EZ or EZ Protect+ _____
- f. **Do not proceed if c., d., and e. are not met** _____

7. Test Procedure _____

- a. Set out required number of test strips and place them in a dry labeled container at room temperature, or take out strips as needed. _____
 1. Discard unused test strips at the end of the day _____
- b. Label test strips, one for each test sample and each control. Avoid crushing sample compartment(s). Re-shape dented sample compartments to fit into incubator. Discard strips that cannot be re-shaped. _____

- c. Mix milk sample(s)/control(s) 25 times in 7 sec with a 1 ft movement or vortex for 10 sec at maximum setting; use within 3 min (samples/controls must be in appropriate containers to allow the use of vortexing) _____
 - 1. Centrifuge sheep milk sample(s)/controls that have been previously frozen; refer to 5.d.2.a-c and 5.e.2.a-c _____
- d. Place strip into appropriate incubator _____
 - 1. EZ reader and EZ Protect+ in incubate and read mode displays appropriate test name and message when in correct temperature range, 55-57°C _____
 - a. EZ reader displays “add milk to strip and close door”. Follow item 7.i _____
 - b. EZ Protect+ displays “add milk to strip”. Follow item 7.i within 1 minute of message display _____
- e. While holding strip flat, peel back plastic (to ‘peel to here’ line) to expose sample pad compartment. Avoid lifting the wick and sponge under tape _____
 - 1. For multiple samples, complete steps 7.d-g for each sample/control, before starting test of next sample _____
 - 2. Complete all samples within 2 min (1 min 15 sec for SL3 test) of placing first strip in incubator _____
- f. Add 300 µL of mixed sample/control to corresponding strip _____
 - 1. Using pipettor (item 3.c) with new tip for each sample/control, draw up 300 µL avoiding foam and bubbles _____
 - a. Remove tip from liquid _____
 - b. While holding the pipettor vertically, expel test portion slowly into either side well of appropriate strip _____
 - 2. Using new manufacturer-provided ROSA-pipet (item 3.d) for each sample/control **[Screening only]** _____
 - a. Squeeze top bulb while holding vertically with bulb and overflow reservoir side pointing down, draw up test portion avoiding foam and bubbles. Sample should completely fill pipet shaft and overflow into the bottom half of the overflow reservoir _____
 - b. Remove tip from liquid _____
 - c. While holding the ROSA-pipet vertically, expel test portion slowly into either side well of appropriate strip. Excess portion should remain in reservoir _____
- g. Re-seal plastic firmly around sample pad compartment. DO NOT press on compartment or sponge in compartment _____

- h. ROSA Reader and Charm EZ and EZ Protect+ (read only mode) _____
1. Close lid and latch ROSA incubator to start automatic timer in the incubator. If no automatic timer in incubator, set external timer for 8 min for SL. For SL test, incubate 8 min not to exceed 9 min. For SL3 test, incubate 3 min not to exceed 3 min and 30 sec _____
 2. At end of incubation visually inspect C (Control) line. An absent C line, a partial C line or an indistinct C line indicates an invalid test; and the sample/control must be re-tested _____
 3. Insert only valid test(s) in reader _____
 - a. ROSA reader set to appropriate channel _____
 1. SLBL slow blink for SL beta-lactam test _____
 2. SLBL solid (no blink) for SL3 beta-lactam test _____
 3. Press ENTER, reading and interpretation appear in 5 sec, read strips within 5 min (3 min with SL3) of completion of incubation _____
 - b. Charm EZ and EZ Protect+ automatically sets channel when strip inserted. Read strips within 5 min (3 min with SL3) of completion of incubation _____
 1. Charm EZ- Close door; reading and interpretation appear in 5 sec, read strips within 5 min (3 min with SL3) of completion of incubation _____
 2. Charm EZ Protect+ – reading and interpretation appear in 5 sec automatically on strip insertion, detection and entering required information _____
- i. Charm EZ and EZ Protect+ (incubate and read mode) _____
1. Charm EZ and EZ Protect+ automatically sets channel and incubator temperature when strip inserted. Optionally enter sample ID and/or operator. Wait for “Add milk to strip” message indicating incubator temperature is in range _____
 2. Peel strip (7.e), add milk (7.f), and re-seal strip (7.g) _____
 3. Start Reading _____
 - a. EZ-Reader - Close door to begin _____
 - b. EZ-Protect+- Count starts automatically on strip and milk flow detection and times out if milk not detected in specified time _____

4. Charm EZ and EZ Protect+ automatically prompts for further testing when positive

8. Interpretation with Reader

a. If there is a negative or zero reading on the reader, sample is a **Negative (NF)**

b. If there is a positive reading on the reader, sample is an **Initial Positive**

9. Verification of Initial Positive Tanker Samples (see App. N GR item 11); Confirmation of Presumptive Positive Tanker Samples (see App. N GR item 12); and Traceback of Producer(s) on a Confirmed Positive Tanker (see App. N GR item 13)

10. Reporting (see App. N GR item 14)
